



Ministerio de Fomento



CENTRO ESPAÑOL DE METROLOGÍA  
C/ del Alfar, nº 2 - 28760 Tres Cantos (Madrid)

Member State  
SPAIN

OIML Certificate N°:  
R60/1991-ES- 99.01

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **Centro Español de Metrología**

Address: **C/ Alfar, 2  
E-28760 Tres Cantos - Madrid (Spain)**

Person responsible: **José A. Robles Carbonell  
Head of force division**

### Applicant

Name: **Transdutec S.A.**

Address: **C/ Joan Miró, 11  
Sant Adrià del Besòs (Barcelona)  
España**

**Manufacturer:** **Transdutec, S.A**  
of the certified pattern:

**Identification:** **Type TCC-1**  
of the certified pattern: **Futher characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern ( represented by the samples identified in the associated test report ) with the requirements of the following Recommendation of the International Organization of Legal Metrology ( OIML ):

**R60**  
**edition 1991 (E)**  
**for accuracy class C**  
**P<sub>1</sub> = 0.7**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

**This certificate does not bestow any formal of legal international approval.**



Ministerio de Fomento



CENTRO ESPAÑOL DE METROLOGÍA  
C/ del Alfar, nº 2 - 28760 Tres Cantos (Madrid)

OIML Certificate N°:  
R60/1991-ES- 99.01

The conformity was established by test described in the associated test report N° 1998-5.1-229 that includes 22 pages

The issuing authority



Centro Español de Metrología  
MINISTERIO DE FOMENTO

Date : January 22, 1999

The CIML member



Ángel García San Román

Director

Date : January 22, 1999

**Characteristics:**

Maximum number of LC verification interval ( $n_{LC}$ )	3000
Maximum capacity ( $E_{max}$ )	300 kg, 500 kg, 750 kg, 1000 kg and 1500 kg
Accuracy class	C
Temperature range	-10/40 °C
Rated output (C)	2 mV/V
Input impedance ( $R_{LC}$ )	386 ohm $\pm 2\%$
Safe overload, relative	125 % $E_{max}$
Maximum excitation voltage	18 V
Ratio of minimum LC verification interval ( $Y = E_{max}/V_{min}$ )	8500
Minimum dead load, relative ( $E_{min}/E_{max}$ )	0 %

**Important note:** Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.